

狀態:	New	開始日期:	2024-04-23
優先權:	Normal	完成日期:	
被分派者:	Chifu Chung	完成百分比:	0%
分類:		預估工時:	0:00 小時
版本:		耗用工時:	0:00 小時

概述

前情提要：12月測試結果過好，觀察測試資料二維光譜

下表結果：一維訊號


測試集1為12月中10%之資料，測試集2為8、9、10月中之資料

模組：A1	測試集1 Accuracy	測試集1 Precision	測試集1 Recall	測試集1 F1-Score	測試集2 Accuracy	測試集2 Precision	測試集2 Recall	測試集2 F1-Score
RBC	100%	100%	100%	100%	60.81%	43.95%	61.81%	51.37%
Na	85.71%	85.33%	96.96%	90.78%	55.70%	65.22%	61.68%	63.40%
GOT	96.70%	94.82%	100%	97.34%	53.72%	61.93%	60.50%	61.21%
PLT	100%	100%	100%	100%	45.35%	59.79%	51.33%	55.24%
Pro	90.10%	100%	88.61%	93.96%	61.28%	75.57%	62.93%	68.67%
P	91.21%	78.38%	100%	87.88%	68.14%	51.40%	76.92%	61.62%
Ca	100%	100%	100%	100%	50.12%	50.79%	73.12%	59.94%


1. 上次測試數據之光譜二維影像

RBC

RBC 12月測試集 全都只有正常
沒有偏高或偏低




GOT




1. 上次測試數據之光譜二維影像

Na




PLT




1. 上次測試數據之光譜二維影像

Pro




Ca



1. 上次測試數據之光譜二維影像

P



檔案

clipboard-202404231738-2ulbj.png

88.3 KB

2024-04-23

Chifu Chung

clipboard-202404231738-k7rz0.png

125 KB

2024-04-23

Chifu Chung

clipboard-202404231738-hdrzz.png	118 KB	2024-04-23	Chifu Chung
clipboard-202404231738-e4s6t.png	123 KB	2024-04-23	Chifu Chung
clipboard-202404231738-w315a.png	63.1 KB	2024-04-23	Chifu Chung